

CERTIFICATE OF ANALYSIS



Product Name: **TestUrGut qPCR Kit**

Number of reactions: **24**

Catalogue Number: **TUG-01-1024**

Lot Number: **0000918121**

Manufacturing Date: **2025/07/17**

Expiration Date: **2027/07/16**

Test parameter	Test description	Acceptance criteria	Result
Accuracy TestUrGut qPCR Kit	Real-Time PCR	The Ct values obtained for the kit's positive controls fall within $\pm 10\%$ of the average value (as specified in the Technical Specifications).	Passed
Sensitivity and specificity of megamixes	Real-Time PCR	An adequate amplification curve is observed when certified external specific PCR products (C+) and DNA samples are analysed. No amplification is detected in the non-template control (NTC), or amplification appears beyond the limit Ct value (as stated in the Technical Specifications).	Passed
Sensitivity and specificity of positive controls (C+)	Real-Time PCR	An adequate amplification curve is observed when the C+ is analysed using an external megamix. No amplification is detected in the NTC, or amplification appears beyond the limit Ct value (as defined in the Technical Specifications).	Passed
Efficiency of Multiplex qPCR Master Mix	Real-Time PCR	The qPCR efficiency for the Eubacteria marker is between 90% and 110%. No amplification is detected in the NTC, or amplification appears beyond the limit Ct value (as specified in the Technical Specifications).	Passed
Visual inspection control	Visual inspection	Packaging, kit components and labels are in accordance with the specifications.	Passed

CERTIFICATE OF ANALYSIS



GOODGUT, S.L.U. manufactures this product for its specifically intended use. Please refer to the product label and accompanying documentation for detailed intended use statements. All GOODGUT S.L.U. products are manufactured in accordance with the the ISO13485:2016 Quality Management System and applicable regulations for *in vitro* medical devices.

Storage

The TestUrGut qPCR Kit must be stored immediately upon receipt at -15 to -30°C in a temperature-controlled freezer.

Certification Statement:

This certifies that the product has been manufactured and tested in accordance with GOODGUT S.L.U. Quality Control (QC) and Quality Assurance (QA) procedures, and has met all applicable GOODGUT, S.L.U. quality requirements.

Date of signature:

Aug 01, 2025

A handwritten signature in purple ink, appearing to read "J. Gómez", written over a faint, large, stylized "G" watermark.

Dr. Juan Vicente Gómez

Quality Assurance Manager



GOODGUT
PARC DE RECERCA I INNOVACIÓ DE LA UDG
Edifici Centre d'Empreses Giroemprèn
c/Pic de Peguera 11 - 17003 Girona
Tel. +34 972 18 32 20
www.goodgut.eu

GOODGUT S.L.U.

CIF/NIF: B55206916

Pic de Peguera, 11 | 17003 Girona, Catalonia, Spain

Phone: +34 972 18 32 20

E-mail: info@goodgut.eu

www.goodgut.eu