

CERTIFICATE OF ANALYSIS

Product Name: **RAID-Dx qPCR Kit**

Number of reactions: **24 reactions**

Catalogue Number: **DX-02-1024-01-LP**

Lot Number: **RAI124L-121013**

Manufacturing Date: **2025/11**

Expiration Date: **2028/11**

Test parameter	Test description	Acceptance criteria	Result
<i>Accuracy</i> RAID-Dx	Real Time PCR	PCR accuracy and Cq values of the positive controls are within the specified range	Passed
<i>Specificity</i> GG1 Positive control, GG2 Positive control, and GG3 Positive control	Real Time PCR	Specific PCR products are present	Passed
<i>Analytical sensitivity test</i> Multiplex GG1 channels FAM, HEX, and ROX	Real Time PCR	Specific PCR products are present, and the yield is above the specified limit	Passed
<i>Analytical sensitivity test</i> Multiplex GG2 channels FAM, HEX, and ROX	Real Time PCR	Specific PCR products are present, and the yield is above the specified limit	Passed
<i>Analytical sensitivity test</i> Multiplex GG3 channels FAM, HEX, and ROX	Real Time PCR	Specific PCR products are present, and the yield is above the specified limit	Passed

GOODGUT S.L.U. manufactures this product for its specific intended use. See product label and directions for intended use statements. GOODGUT S.L.U. products are manufactured in accordance with the ISO13485:2016 Quality Management System and the current regulations on In Vitro Medical Devices.

Storage

The RAID-Dx should be stored at a constant room temperature 2 to 40 °C. It is recommended for constant storage at a temperature of 25° +/- 5.

CERTIFICATE OF ANALYSIS

Certification Statement:

This certifies that the product was manufactured and tested in accordance with GOODGUT, S.L.U QC and QA procedures. The product has met all GOODGUT S.L.U. quality requirements.

Date of signature:

Jan 14, 2026



Dr. Juan Vicente Gómez

Quality Assurance Manager



GOODGUT
Enhancing digestive health

GOODGUT S.L.U

CIF/NIF: B55206916

Pic de Peguera, 11 | 17003 Girona, Cataluña, España

Teléfono: +34 972 18 32 20

E-mail: info@goodgut.eu

www.goodgut.eu



GOODGUT
Enhancing digestive health

PARC CIENTÍFIC I TECNOLÒGIC UDG
EDIFICI CENTRE D'EMPRESSES
C/ Pic de Peguera, 11 - 17003 Girona
Tel. +34 972 18 32 20
www.goodgut.eu